

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.313 °C	31.2 °C	0.887 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.836 %HR	30 %HR	-0.836 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.732 %HR	46 %HR	-3.732 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	69.803 %HR	66 %HR	-3.803 %HR	± 5.0 %HR	± 1.9E+00 %HR