

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	32.108 °C	32.0 °C	-0.108 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	28.740 %HR	32.4 %HR	3.660 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.325 %HR	49.3 %HR	-1.025 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	71.450 %HR	67.5 %HR	-3.950 %HR	± 5.0 %HR	± 1.9E+00 %HR