

| Nominal ( <i>Nominal</i> ) | Patrón ( <i>Standard</i> ) | UBP ( <i>UUT</i> ) | Error ( <i>Error</i> ) | Exactitud ( <i>Accuracy</i> ) | Incertidumbre ( <i>Uncertainty</i> ) |
|----------------------------|----------------------------|--------------------|------------------------|-------------------------------|--------------------------------------|
| 15 °C                      | 14.951 °C                  | 14.8 °C            | -0.151 °C              | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 25 °C                      | 25.189 °C                  | 24.9 °C            | -0.289 °C              | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 32 °C                      | 32.108 °C                  | 31.9 °C            | -0.208 °C              | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 20 %HR                     | 20.560 %HR                 | 24.9 %HR           | 4.340 %HR              | ± 5.0 %HR                     | ± 1.9E+00 %HR                        |
| 50 %HR                     | 50.325 %HR                 | 50.1 %HR           | -0.225 %HR             | ± 5.0 %HR                     | ± 1.9E+00 %HR                        |
| 70 %HR                     | 71.450 %HR                 | 69.2 %HR           | -2.250 %HR             | ± 5.0 %HR                     | ± 1.9E+00 %HR                        |