

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
15 °C	14.951 °C	15.2 °C	0.249 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.189 °C	25.0 °C	-0.189 °C	± 0.5 °C	± 4.0E-01 °C
32 °C	32.108 °C	31.7 °C	-0.408 °C	± 0.5 °C	± 4.0E-01 °C
20 %HR	20.560 %HR	24.9 %HR	4.340 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.325 %HR	50.0 %HR	-0.325 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	71.450 %HR	69.7 %HR	-1.750 %HR	± 5.0 %HR	± 1.9E+00 %HR