

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
15 °C	15.071 °C	14.8 °C	-0.271 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.359 °C	25.1 °C	-0.259 °C	± 0.5 °C	± 4.0E-01 °C
32 °C	31.887 °C	31.8 °C	-0.087 °C	± 0.5 °C	± 4.0E-01 °C
20 %HR	21.480 %HR	25.1 %HR	3.620 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	51.347 %HR	53.1 %HR	1.753 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	71.081 %HR	67.4 %HR	-3.681 %HR	± 5.0 %HR	± 1.9E+00 %HR