

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
15 °C	15.071 °C	15.1 °C	0.029 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.359 °C	25.3 °C	-0.059 °C	± 0.5 °C	± 4.0E-01 °C
32 °C	31.887 °C	31.6 °C	-0.287 °C	± 0.5 °C	± 4.0E-01 °C
20 %HR	21.480 %HR	24.5 %HR	3.020 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	51.347 %HR	53.1 %HR	1.753 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	71.081 %HR	68.4 %HR	-2.681 %HR	± 5.0 %HR	± 1.9E+00 %HR