

Clase de Exactitud (Accuracy Class): Clase III (Class III)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre
m	m	mm	mm	(Uncertainty)	m	m	mm	mm	(Uncertainty)
0 a 0.1	0.09920	0.800	± 1.2	± 4.3E-01 mm					
0 a 0.2	0.19931	0.690	± 1.3	± 4.3E-01 mm					
0 a 0.3	0.29924	0.760	± 1.4	± 4.3E-01 mm					
0 a 0.4	0.39928	0.720	± 1.5	± 4.3E-01 mm					
0 a 0.5	0.49915	0.850	± 1.6	± 4.3E-01 mm					
0 a 0.6	0.59915	0.850	± 1.6	± 4.3E-01 mm					
0 a 0.7	0.69912	0.880	± 1.7	± 4.3E-01 mm					
0 a 0.8	0.79908	0.920	± 1.8	± 4.3E-01 mm					
0 a 0.9	0.89908	0.920	± 1.9	± 4.3E-01 mm					
0 a 1	0.99904	0.960	± 2	± 4.3E-01 mm					
0 a 2	1.99978	0.930	± 2.8	± 4.8E-01 mm					
0 a 3	2.99977	0.900	± 3.6	± 4.8E-01 mm					
0 a 4	3.99980	0.820	± 4.4	± 4.8E-01 mm					
0 a 5	4.99933	0.666	± 5.2	± 4.8E-01 mm					
0 a 6	5.99925	0.754	± 6	± 3.6E-01 mm					
0 a 7	6.99932	0.682	± 6.8	± 3.6E-01 mm					
0 a 8	7.99921	0.791	± 7.6	± 3.6E-01 mm					