

Tipo de Medida (<i>Measurement Type</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	E..M.P (<i>Tolerance</i>)	Incertidumbre (<i>Uncertainty</i>)
Exterior / (Exterior)	4.00 in	4.002 in	0.002 in	± 0.0028 in	± 1.3E-03 in
Exterior / (Exterior)	4.80 in	4.801 in	0.001 in	± 0.0028 in	± 1.3E-03 in
Exterior / (Exterior)	6.40 in	6.401 in	0.001 in	± 0.0028 in	± 1.3E-03 in
Exterior / (Exterior)	7.20 in	7.202 in	0.002 in	± 0.0028 in	± 1.3E-03 in
Exterior / (Exterior)	8.00 in	8.002 in	0.002 in	± 0.0031 in	± 1.3E-03 in
Interior / (Interior)	0.80 in	0.800 in	0.000 in	± 0.0020 in	± 7.4E-04 in
Interior / (Interior)	7.20 in	7.201 in	0.001 in	± 0.0028 in	± 7.4E-04 in
Profundidad / (Depth)	0.80 in	0.802 in	0.002 in	± 0.0020 in	± 8.7E-04 in
Profundidad / (Depth)	7.20 in	7.202 in	0.002 in	± 0.0028 in	± 8.7E-04 in