

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.313 °C	30.2 °C	-0.113 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	31.858 %HR	34.2 %HR	2.342 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	49.732 %HR	52.2 %HR	2.468 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	69.803 %HR	72.8 %HR	2.997 %HR	± 3.0 %HR	± 1.9E+00 %HR