

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.313 °C	30.2 °C	-0.113 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	31.858 %HR	32.6 %HR	0.742 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	49.732 %HR	51.7 %HR	1.968 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	69.803 %HR	72.1 %HR	2.297 %HR	± 3.0 %HR	± 1.9E+00 %HR