

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
1.00 mm	1.000 mm	0 µm	± 30 µm	±	5.8E-01 µm
1.25 mm	1.250 mm	0 µm	± 30 µm	±	5.8E-01 µm
1.50 mm	1.500 mm	0 µm	± 30 µm	±	5.8E-01 µm
2.00 mm	2.002 mm	2 µm	± 30 µm	±	5.8E-01 µm
3.00 mm	3.003 mm	3 µm	± 30 µm	±	5.8E-01 µm
5.00 mm	5.003 mm	3 µm	± 30 µm	±	5.8E-01 µm
8.00 mm	8.003 mm	3 µm	± 30 µm	±	5.9E-01 µm
10.00 mm	10.003 mm	3 µm	± 30 µm	±	5.9E-01 µm
12.50 mm	12.503 mm	3 µm	± 30 µm	±	5.9E-01 µm