

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
1.00 mm	1.000 mm	0 μm	± 30 μm	±	5.8E-01 μm
1.25 mm	1.250 mm	0 μm	± 30 μm	±	5.8E-01 μm
1.50 mm	1.501 mm	1 μm	± 30 μm	±	5.8E-01 μm
2.00 mm	2.001 mm	1 μm	± 30 μm	±	5.8E-01 μm
3.00 mm	3.001 mm	1 μm	± 30 μm	±	5.8E-01 μm
5.00 mm	5.001 mm	1 μm	± 30 μm	±	5.8E-01 μm
8.00 mm	8.002 mm	2 μm	± 30 μm	±	5.9E-01 μm
10.00 mm	10.002 mm	2 μm	± 30 μm	±	5.9E-01 μm
12.50 mm	12.501 mm	1 μm	± 30 μm	±	5.9E-01 μm