

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.293 °C	29.7 °C	-0.593 °C	± 1.0 °C	± 4.0E-01 °C
35 %HR	34.793 %HR	36 %HR	1.207 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	52.666 %HR	50 %HR	-2.666 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.263 %HR	66 %HR	-4.263 %HR	± 5.0 %HR	± 1.9E+00 %HR