

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.293 °C	30.5 °C	0.207 °C	± 1.0 °C	± 4.0E-01 °C
35 %HR	34.793 %HR	39 %HR	4.207 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	52.666 %HR	56 %HR	3.334 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.263 %HR	73 %HR	2.737 %HR	± 5.0 %HR	± 1.9E+00 %HR