

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.293 °C	29.5 °C	-0.793 °C	± 1.0 °C	± 4.0E-01 °C
35 %HR	34.793 %HR	41 %HR	6.207 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	52.666 %HR	64 %HR	11.334 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.263 %HR	84 %HR	13.737 %HR	± 5.0 %HR	± 1.9E+00 %HR