

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.082 °C	28.8 °C	-1.282 °C	± 1.0 °C	± 4.0E-01 °C
35 %HR	34.916 %HR	36 %HR	1.084 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	51.756 %HR	56 %HR	4.244 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	69.537 %HR	75 %HR	5.463 %HR	± 5.0 %HR	± 1.9E+00 %HR