

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.293 °C	29.9 °C	-0.393 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	29.680 %HR	28.6 %HR	-1.080 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	52.666 %HR	50.0 %HR	-2.666 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	75.376 %HR	73.0 %HR	-2.376 %HR	± 3.0 %HR	± 1.9E+00 %HR