

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
4 °C	3.810 °C	3.7 °C	-0.110 °C	± 0.5 °C	± 4.0E-01 °C
15 °C	14.901 °C	14.9 °C	-0.001 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.082 °C	29.6 °C	-0.482 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.826 %HR	28.1 %HR	-2.726 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	51.756 %HR	47.4 %HR	-4.356 %HR	± 3.0 %HR	± 1.9E+00 %HR
65 %HR	66.470 %HR	59.9 %HR	-6.570 %HR	± 3.0 %HR	± 1.9E+00 %HR