

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
4 °C	3.810 °C	3.7 °C	-0.110 °C	± 0.5 °C	± 4.0E-01 °C
15 °C	14.901 °C	15.0 °C	0.099 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.082 °C	30.0 °C	-0.082 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.826 %HR	29.6 %HR	-1.226 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	51.756 %HR	49.3 %HR	-2.456 %HR	± 3.0 %HR	± 1.9E+00 %HR
65 %HR	66.470 %HR	63.2 %HR	-3.270 %HR	± 3.0 %HR	± 1.9E+00 %HR