

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
4 °C	3.810 °C	2.7 °C	-1.110 °C	± 0.5 °C	± 4.0E-01 °C
15 °C	14.901 °C	13.8 °C	-1.101 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.082 °C	28.8 °C	-1.282 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.826 %HR	28.4 %HR	-2.426 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	51.756 %HR	47.5 %HR	-4.256 %HR	± 3.0 %HR	± 1.9E+00 %HR
65 %HR	66.470 %HR	60.9 %HR	-5.570 %HR	± 3.0 %HR	± 1.9E+00 %HR