

Certificado No. (Certificate #):		54241		Equipo (Instrument):		Calibrador Pie de Rey	
Fecha de Calibración (Calibration Date):		2025-02-21		Marca (Brand):		Mitutoyo	
Tipo de Medida (Measurement Type)		Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
Exterior / (Exterior)		1 mm	1.00 mm	0 µm	± 30 µm	±	7.4E+00 µm
Exterior / (Exterior)		15 mm	15.00 mm	0 µm	± 30 µm	±	7.4E+00 µm
Exterior / (Exterior)		50 mm	50.00 mm	0 µm	± 30 µm	±	7.4E+00 µm
Exterior / (Exterior)		135 mm	135.00 mm	0 µm	± 30 µm	±	7.4E+00 µm
Exterior / (Exterior)		150 mm	150.00 mm	0 µm	± 30 µm	±	7.4E+00 µm
Interior / (Interior)		20 mm	20.00 mm	0 µm	± 30 µm	±	8.2E+00 µm
Interior / (Interior)		100 mm	100.00 mm	0 µm	± 30 µm	±	8.2E+00 µm
Profundidad / (Depth)		30 mm	30.00 mm	0 µm	± 30 µm	±	7.1E+00 µm
Profundidad / (Depth)		50 mm	50.00 mm	0 µm	± 30 µm	±	7.1E+00 µm