

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.190 °C	30.5 °C	0.310 °C	± 0.5 °C	± 4.0E-01 °C
35 %HR	33.433 %HR	32 %HR	-1.433 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.077 %HR	48 %HR	-1.077 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.928 %HR	69 %HR	-1.928 %HR	± 5.0 %HR	± 1.9E+00 %HR