

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
35 °C	35.192 °C	35.7 °C	0.508 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	28.689 %HR	29.8 %HR	1.111 %HR	± 3 %HR	± 1.9E+00 %HR
50 %HR	50.693 %HR	49.2 %HR	-1.493 %HR	± 3 %HR	± 1.9E+00 %HR
70 %HR	69.466 %HR	67.5 %HR	-1.966 %HR	± 3 %HR	± 1.9E+00 %HR