

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
20 °C	20.025 °C	20.0 °C	-0.025 °C	± 0.5 °C	± 4.0E-01 °C
35 %HR	33.535 %HR	35.5 %HR	1.965 %HR	± 1.1 %HR	± 1.9E+00 %HR
50 %HR	49.589 %HR	51.4 %HR	1.811 %HR	± 1.5 %HR	± 1.9E+00 %HR
70 %HR	70.448 %HR	72.3 %HR	1.852 %HR	± 2.1 %HR	± 1.9E+00 %HR