

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-5 °C	-5.235 °C	-6.0 °C	-0.765 °C	± 1.0 °C	± 4.0E-01 °C
0 °C	-0.201 °C	-1 °C	-0.799 °C	± 1.0 °C	± 4.0E-01 °C
10 °C	9.807 °C	9 °C	-0.807 °C	± 1.0 °C	± 4.0E-01 °C
20 °C	19.834 °C	19 °C	-0.834 °C	± 1.0 °C	± 4.0E-01 °C
40 °C	39.990 °C	39.0 °C	-0.990 °C	± 1.0 °C	± 4.0E-01 °C