

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
10 °C	10.809 °C	10.3 °C	-0.509 °C	± 3.0 °C	± 1.8E-01 °C
20 °C	19.834 °C	19.6 °C	-0.234 °C	± 3.0 °C	± 1.8E-01 °C
30 °C	30.062 °C	29.9 °C	-0.162 °C	± 3.0 °C	± 1.6E-01 °C
30 %HR	30.059 %HR	32.3 %HR	2.241 %HR	± 3.0 %HR	± 1.0E+00 %HR
40 %HR	40.284 %HR	42.8 %HR	2.516 %HR	± 3.0 %HR	± 1.0E+00 %HR
50 %HR	49.998 %HR	52.2 %HR	2.202 %HR	± 3.0 %HR	± 1.2E+00 %HR
70 %HR	70.448 %HR	68.1 %HR	-2.348 %HR	± 3.0 %HR	± 1.2E+00 %HR
80 %HR	80.540 %HR	78.3 %HR	-2.240 %HR	± 3.0 %HR	± 1.1E+00 %HR