

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
10 °C	10.809 °C	10.2 °C	-0.609 °C	± 3.0 °C	± 1.8E-01 °C
20 °C	19.834 °C	19.5 °C	-0.334 °C	± 3.0 °C	± 1.8E-01 °C
30 °C	30.062 °C	29.7 °C	-0.362 °C	± 3.0 °C	± 1.6E-01 °C
30 %HR	30.059 %HR	32.4 %HR	2.341 %HR	± 3.0 %HR	± 1.0E+00 %HR
40 %HR	40.284 %HR	42.5 %HR	2.216 %HR	± 3.0 %HR	± 1.0E+00 %HR
50 %HR	49.998 %HR	52.6 %HR	2.602 %HR	± 3.0 %HR	± 1.2E+00 %HR
70 %HR	70.448 %HR	71.7 %HR	1.252 %HR	± 3.0 %HR	± 1.2E+00 %HR
80 %HR	80.540 %HR	77.9 %HR	-2.640 %HR	± 3.0 %HR	± 1.1E+00 %HR