

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
10 °C	10.809 °C	10.5 °C	-0.309 °C	± 3.0 °C	± 1.8E-01 °C
20 °C	19.834 °C	19.7 °C	-0.134 °C	± 3.0 °C	± 1.8E-01 °C
30 °C	30.062 °C	29.9 °C	-0.162 °C	± 3.0 °C	± 1.6E-01 °C
30 %HR	30.059 %HR	32.4 %HR	2.341 %HR	± 3.0 %HR	± 1.0E+00 %HR
40 %HR	40.284 %HR	40.8 %HR	0.516 %HR	± 3.0 %HR	± 1.0E+00 %HR
50 %HR	49.998 %HR	49.6 %HR	-0.398 %HR	± 3.0 %HR	± 1.2E+00 %HR
70 %HR	70.448 %HR	68.3 %HR	-2.148 %HR	± 3.0 %HR	± 1.2E+00 %HR
80 %HR	80.540 %HR	77.8 %HR	-2.740 %HR	± 3.0 %HR	± 1.1E+00 %HR