

| Nominal (Nominal) | Patrón (Standard) | UBP (UUT) | Error (Error) | Exactitud (Accuracy) | Incertidumbre (Uncertainty) |
|-------------------|-------------------|-----------|---------------|----------------------|-----------------------------|
| 10 °C             | 10.809 °C         | 11.3 °C   | 0.491 °C      | ± 3.0 °C             | ± 1.8E-01 °C                |
| 20 °C             | 19.834 °C         | 20.5 °C   | 0.666 °C      | ± 3.0 °C             | ± 1.8E-01 °C                |
| 30 °C             | 30.062 °C         | 30.6 °C   | 0.538 °C      | ± 3.0 °C             | ± 1.6E-01 °C                |
| 30 %HR            | 30.059 %HR        | 35.3 %HR  | 5.241 %HR     | ± 3.0 %HR            | ± 1.0E+00 %HR               |
| 40 %HR            | 40.284 %HR        | 45.2 %HR  | 4.916 %HR     | ± 3.0 %HR            | ± 1.0E+00 %HR               |
| 50 %HR            | 49.998 %HR        | 54.3 %HR  | 4.302 %HR     | ± 3.0 %HR            | ± 1.2E+00 %HR               |
| 70 %HR            | 70.448 %HR        | 71.5 %HR  | 1.052 %HR     | ± 3.0 %HR            | ± 1.2E+00 %HR               |
| 80 %HR            | 80.540 %HR        | 80.4 %HR  | -0.140 %HR    | ± 3.0 %HR            | ± 1.1E+00 %HR               |