

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-20 °C	-20.276 °C	-19.9 °C	0.376 °C	± 3.0 °C	± 2.2E-01 °C
0 °C	-0.221 °C	0.3 °C	0.521 °C	± 3.0 °C	± 2.2E-01 °C
20 °C	20.035 °C	19.3 °C	-0.735 °C	± 3.0 °C	± 1.8E-01 °C
40 °C	40.391 °C	40.6 °C	0.209 °C	± 3.0 °C	± 1.6E-01 °C
60 °C	60.446 °C	60.9 °C	0.454 °C	± 3.0 °C	± 1.1E+00 °C
30 %HR	30.059 %HR	32.6 %HR	2.541 %HR	± 3.0 %HR	± 1.0E+00 %HR
40 %HR	40.284 %HR	42.9 %HR	2.616 %HR	± 3.0 %HR	± 1.0E+00 %HR
50 %HR	49.998 %HR	52.7 %HR	2.702 %HR	± 3.0 %HR	± 1.2E+00 %HR
70 %HR	70.448 %HR	68.9 %HR	-1.548 %HR	± 3.0 %HR	± 1.2E+00 %HR
80 %HR	80.540 %HR	78.2 %HR	-2.340 %HR	± 3.0 %HR	± 1.1E+00 %HR