

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-20 °C	-20.276 °C	-19.4 °C	0.876 °C	± 3.0 °C	± 2.2E-01 °C
0 °C	-0.221 °C	-0.6 °C	-0.379 °C	± 3.0 °C	± 2.2E-01 °C
20 °C	20.035 °C	19.7 °C	-0.335 °C	± 3.0 °C	± 1.8E-01 °C
40 °C	40.391 °C	40.9 °C	0.509 °C	± 3.0 °C	± 1.6E-01 °C
60 °C	60.446 °C	60.8 °C	0.354 °C	± 3.0 °C	± 1.1E+00 °C
30 %HR	30.059 %HR	32.9 %HR	2.841 %HR	± 3.0 %HR	± 1.0E+00 %HR
40 %HR	40.284 %HR	42.8 %HR	2.516 %HR	± 3.0 %HR	± 1.0E+00 %HR
50 %HR	49.998 %HR	52.4 %HR	2.402 %HR	± 3.0 %HR	± 1.2E+00 %HR
70 %HR	70.448 %HR	68.1 %HR	-2.348 %HR	± 3.0 %HR	± 1.2E+00 %HR
80 %HR	80.540 %HR	78.3 %HR	-2.240 %HR	± 3.0 %HR	± 1.1E+00 %HR