

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-20 °C	-20.276 °C	-21.0 °C	-0.724 °C	± 3.0 °C	± 2.2E-01 °C
0 °C	-0.221 °C	0.1 °C	0.321 °C	± 3.0 °C	± 2.2E-01 °C
20 °C	20.035 °C	20.2 °C	0.165 °C	± 3.0 °C	± 1.8E-01 °C
40 °C	40.391 °C	40.2 °C	-0.191 °C	± 3.0 °C	± 1.6E-01 °C
60 °C	60.446 °C	60.7 °C	0.254 °C	± 3.0 °C	± 1.1E+00 °C
30 %HR	30.059 %HR	32.2 %HR	2.141 %HR	± 3.0 %HR	± 1.0E+00 %HR
40 %HR	40.284 %HR	42.7 %HR	2.416 %HR	± 3.0 %HR	± 1.0E+00 %HR
50 %HR	49.998 %HR	52.1 %HR	2.102 %HR	± 3.0 %HR	± 1.2E+00 %HR
70 %HR	70.448 %HR	69.6 %HR	-0.848 %HR	± 3.0 %HR	± 1.2E+00 %HR
80 %HR	80.540 %HR	81.5 %HR	0.960 %HR	± 3.0 %HR	± 1.1E+00 %HR