

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
15 °C	14.696 °C	13.6 °C	-1.096 °C	± 2 °C	± 4.0E-01 °C
25 °C	25.008 °C	23.5 °C	-1.508 °C	± 2 °C	± 4.0E-01 °C
50 °C	50.198 °C	48.1 °C	-2.098 °C	± 2 °C	± 1.1E+00 °C
25 %HR	24.128 %HR	21 %HR	-3.128 %HR	± 5 %HR	± 1.9E+00 %HR
50 %HR	51.429 %HR	50 %HR	-1.429 %HR	± 5 %HR	± 1.9E+00 %HR
75 %HR	72.994 %HR	70 %HR	-2.994 %HR	± 5 %HR	± 1.9E+00 %HR