

| Nominal ( <i>Nominal</i> ) | Patrón ( <i>Standard</i> ) | UBP ( <i>UUT</i> ) | Error ( <i>Error</i> ) | Exactitud ( <i>Accuracy</i> ) |       | Incertidumbre ( <i>Uncertainty</i> ) |             |
|----------------------------|----------------------------|--------------------|------------------------|-------------------------------|-------|--------------------------------------|-------------|
| 15 °C                      | 14.696 °C                  | 14.2 °C            | -0.496 °C              | ±                             | 2 °C  | ±                                    | 4.0E-01 °C  |
| 25 °C                      | 25.008 °C                  | 24.7 °C            | -0.308 °C              | ±                             | 2 °C  | ±                                    | 4.0E-01 °C  |
| 50 °C                      | 50.198 °C                  | 49.8 °C            | -0.398 °C              | ±                             | 2 °C  | ±                                    | 1.1E+00 °C  |
| 25 %HR                     | 24.128 %HR                 | 21 %HR             | -3.128 %HR             | ±                             | 5 %HR | ±                                    | 1.9E+00 %HR |
| 50 %HR                     | 51.429 %HR                 | 50 %HR             | -1.429 %HR             | ±                             | 5 %HR | ±                                    | 1.9E+00 %HR |
| 75 %HR                     | 72.994 %HR                 | 70 %HR             | -2.994 %HR             | ±                             | 5 %HR | ±                                    | 1.9E+00 %HR |