

Certificado No.: 42778

Equipo (Instrument): Termohigrómetro

Fecha de Calibración: 2024-02-22

Marca (Brand): RADIO SHACK

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
# -30 °C	-30.625 °C	-31.4 °C	-0.775 °C	± 1.0 °C	± 0.4 °C
-25 °C	-25.651 °C	-26 °C	-0.649 °C	± 1.0 °C	± 0.4 °C
65 %HR	66.276 %HR	65 %HR	-1.576 %HR	± 3.0 %HR	± 1.9 %HR
70 %HR	70.959 %HR	69 %HR	-1.959 %HR	± 3.0 %HR	± 1.9 %HR