

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
# -30 °C	-30.625 °C	-31.4 °C	-0.775 °C	± 1.0 °C	± 4.0E-01 °C
-25 °C	-25.651 °C	-26.3 °C	-0.649 °C	± 0.5 °C	± 4.0E-01 °C
65 %HR	66.276 %HR	64.7 %HR	-1.576 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.959 %HR	69.0 %HR	-1.959 %HR	± 3.0 %HR	± 1.9E+00 %HR