

Certificado No. (Certificate #):		54376	Equipo (Instrument): Calibrador Pie de Rey		
Fecha de Calibración (Calibration Date):		2025-02-13	Marca (Brand): Performance Tool		
Tipo de Medida (Measurement Type)	Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)
Exterior / (Exterior)	1 mm	0.99 mm	-10 µm	± 30 µm	± 7.4E+00 µm
Exterior / (Exterior)	10 mm	9.99 mm	-10 µm	± 30 µm	± 7.4E+00 µm
Exterior / (Exterior)	25 mm	25.00 mm	0 µm	± 30 µm	± 7.4E+00 µm
Exterior / (Exterior)	50 mm	50.01 mm	10 µm	± 30 µm	± 7.4E+00 µm
Exterior / (Exterior)	150 mm	150.02 mm	20 µm	± 30 µm	± 7.5E+00 µm
Interior / (Interior)	50 mm	49.98 mm	-17 µm	± 30 µm	± 8.2E+00 µm
Interior / (Interior)	100 mm	99.97 mm	-27 µm	± 30 µm	± 8.2E+00 µm
Profundidad / (Depth)	50 mm	50.02 mm	17 µm	± 30 µm	± 7.1E+00 µm
Profundidad / (Depth)	100 mm	100.02 mm	17 µm	± 30 µm	± 7.1E+00 µm