

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
20 °C	20.035 °C	19.3 °C	-0.735 °C	± 1.0 °C	± 4.0E-01 °C
30 °C	30.163 °C	29 °C	-0.963 °C	± 1.0 °C	± 4.0E-01 °C
60 %HR	60.447 %HR	65 %HR	4.553 %HR	± 5.0 %HR	± 1.9E+00 %HR
75 %HR	75.560 %HR	76 %HR	0.140 %HR	± 5.0 %HR	± 1.9E+00 %HR