

Clase de Exactitud (Accuracy Class): Clase III (Class III)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre
m	m	mm	mm	(Uncertainty)	m	m	mm	mm	(Uncertainty)
0 a 0.1	0.09919	0.810	± 1.2	± 4.3E-01 mm					
0 a 0.3	0.29918	0.820	± 1.4	± 4.3E-01 mm					
0 a 0.4	0.39917	0.830	± 1.5	± 4.3E-01 mm					
0 a 0.5	0.49915	0.850	± 1.6	± 4.3E-01 mm					
0 a 0.8	0.79915	0.850	± 1.8	± 4.3E-01 mm					
0 a 0.9	0.89922	0.780	± 1.9	± 4.3E-01 mm					
0 a 1	0.99919	0.810	± 2	± 4.3E-01 mm					
0 a 1.5	1.49916	0.840	± 2.4	± 4.8E-01 mm					
0 a 2	1.99894	1.060	± 2.8	± 4.8E-01 mm					
0 a 2.5	1.57000	1.570	± 3.2	± 4.8E-01 mm					
0 a 3	3.00000	0.660	± 3.6	± 4.8E-01 mm					