

Clase de Exactitud (Accuracy Class): Clase III (Class III)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre
m	m	mm	mm	(Uncertainty)	m	m	mm	mm	(Uncertainty)
0 a 0.75	0.74925	1.020	± 1.8	± 2.9E+00 mm					
0 a 0.85	0.84915	1.120	± 1.8	± 2.9E+00 mm					
0 a 0.95	0.94905	1.230	± 1.9	± 2.9E+00 mm					
0 a 1.05	1.04895	1.330	± 2	± 2.9E+00 mm					
0 a 1.15	1.14885	1.440	± 2.1	± 2.9E+00 mm					
0 a 1.25	1.24875	1.540	± 2.2	± 2.9E+00 mm					
0 a 1.35	1.34865	1.650	± 2.2	± 2.9E+00 mm					
0 a 1.45	1.44855	1.750	± 2.3	± 2.9E+00 mm					
0 a 1.55	1.54845	1.860	± 2.4	± 2.9E+00 mm					
0 a 1.65	1.64835	1.960	± 2.5	± 2.9E+00 mm					
0 a 1.75	1.74825	2.070	± 2.6	± 2.9E+00 mm					
0 a 1.85	1.84815	2.170	± 2.6	± 2.9E+00 mm					
0 a 1.95	1.94805	2.280	± 2.7	± 2.9E+00 mm					