

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.163 °C	29.8 °C	-0.363 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.366 %HR	35 %HR	4.634 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.989 %HR	52 %HR	1.011 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.959 %HR	74 %HR	3.041 %HR	± 5.0 %HR	± 1.9E+00 %HR