

Clase de Exactitud (Accuracy Class): Clase II (Class II)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre
m	m	mm	mm	(Uncertainty)	m	m	mm	mm	(Uncertainty)
0 a 0.8	0.79975	0.520	± 0.9	± 1.5E+00 mm					
0 a 0.9	0.89975	0.520	± 0.9	± 1.5E+00 mm					
0 a 1	0.99971	0.570	± 1	± 1.5E+00 mm					
0 a 1.1	1.09970	0.590	± 1	± 1.5E+00 mm					
0 a 1.2	1.19969	0.600	± 1	± 1.5E+00 mm					
0 a 1.3	1.29957	0.730	± 1.1	± 1.5E+00 mm					
0 a 1.4	1.39956	0.740	± 1.1	± 1.5E+00 mm					
0 a 1.5	1.49946	0.840	± 1.2	± 1.5E+00 mm					
0 a 1.6	1.59945	0.860	± 1.2	± 1.5E+00 mm					
0 a 1.7	1.69945	0.860	± 1.2	± 1.5E+00 mm					
0 a 1.8	1.79942	0.900	± 1.3	± 1.5E+00 mm					
0 a 1.9	1.89937	0.960	± 1.3	± 1.5E+00 mm					
0 a 2	1.99934	0.990	± 1.4	± 1.5E+00 mm					
0 a 2.1	2.09931	1.030	± 1.4	± 1.5E+00 mm					