

| Nominal (Nominal) | Patrón (Standard) | UBP (UUT) | Error (Error) | Exactitud (Accuracy) | Incertidumbre (Uncertainty) |
|-------------------|-------------------|-----------|---------------|----------------------|-----------------------------|
| 15 °C             | 14.921 °C         | 14.6 °C   | -0.321 °C     | ± 0.5 °C             | ± 4.0E-01 °C                |
| 25 °C             | 25.049 °C         | 25 °C     | -0.249 °C     | ± 0.5 °C             | ± 4.0E-01 °C                |
| 32 °C             | 32.369 °C         | 32 °C     | -0.469 °C     | ± 0.5 °C             | ± 4.0E-01 °C                |
| 30 %HR            | 30.570 %HR        | 33 %HR    | 2.530 %HR     | ± 3.0 %HR            | ± 1.9E+00 %HR               |
| 70 %HR            | 70.448 %HR        | 73.5 %HR  | 3.052 %HR     | ± 3.0 %HR            | ± 1.9E+00 %HR               |