

Clase de Exactitud (Accuracy Class): Clase II (Class II)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre
m	m	mm	mm	(Uncertainty)	m	m	mm	mm	(Uncertainty)
0 a 0.1	0.09975	0.250	± 0.6	± 1.5E+00 mm					
0 a 0.4	0.39948	0.520	± 0.7	± 1.5E+00 mm					
0 a 0.7	0.69982	0.180	± 0.8	± 1.5E+00 mm					
0 a 1	0.99958	0.420	± 1	± 1.5E+00 mm					
0 a 1.3	1.29965	0.350	± 1.1	± 1.5E+00 mm					
0 a 1.6	1.59951	0.490	± 1.2	± 1.5E+00 mm					
0 a 1.9	1.89925	0.750	± 1.3	± 1.5E+00 mm					
0 a 2.1	2.09933	0.670	± 1.4	± 1.5E+00 mm					
0 a 2.4	2.39943	0.570	± 1.5	± 1.5E+00 mm					
0 a 2.7	2.69899	1.010	± 1.6	± 1.5E+00 mm					
0 a 3	2.99904	0.960	± 1.8	± 1.5E+00 mm					