

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
10 °C	10.308 °C	9.5 °C	-0.808 °C	± 0.5 °C	± 4.0E-01 °C
32 °C	31.867 °C	31.3 °C	-0.567 °C	± 0.5 °C	± 4.0E-01 °C
40 %HR	39.773 %HR	42.8 %HR	3.027 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	75.560 %HR	78.2 %HR	2.640 %HR	± 3.0 %HR	± 1.9E+00 %HR