

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
10 °C	10.308 °C	8.7 °C	-1.608 °C	± 0.5 °C	± 4.0E-01 °C
32 °C	31.867 °C	32.1 °C	0.233 °C	± 0.5 °C	± 4.0E-01 °C
40 %HR	39.773 %HR	43.2 %HR	3.427 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	75.560 %HR	79.2 %HR	3.640 %HR	± 3.0 %HR	± 1.9E+00 %HR