

| Nominal (Nominal) | Patrón (Standard) | UBP (UUT) | Error (Error) | Exactitud (Accuracy) | Incertidumbre (Uncertainty) |
|-------------------|-------------------|-----------|---------------|----------------------|-----------------------------|
| 10 °C             | 10.308 °C         | 10.5 °C   | 0.192 °C      | ± 0.5 °C             | ± 4.0E-01 °C                |
| 32 °C             | 31.867 °C         | 33.0 °C   | 1.133 °C      | ± 0.5 °C             | ± 4.0E-01 °C                |
| 40 %HR            | 39.773 %HR        | 41.0 %HR  | 1.227 %HR     | ± 3.0 %HR            | ± 1.9E+00 %HR               |
| 75 %HR            | 75.560 %HR        | 80.5 %HR  | 4.940 %HR     | ± 3.0 %HR            | ± 1.9E+00 %HR               |