

Clase de Exactitud (Accuracy Class): Clase III (Class III)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)		Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)	
m	m	mm	mm			m	m	mm	mm		
0 a 0.1	0.09917	0.830	± 1.2	±	#REF! mm						
0 a 0.5	0.49918	0.820	± 1.6	±	#REF! mm						
0 a 1	0.99915	0.850	± 2	±	#REF! mm						
0 a 2	1.99915	0.850	± 2.8	±	#REF! mm						
0 a 3	2.99847	1.530	± 3.6	±	#REF! mm						
0 a 4	3.99789	2.110	± 4.4	±	#REF! mm						
0 a 5	5.00128	-1.280	± 5.2	±	#REF! mm						
0 a 6	6.00221	-2.210	± 6	±	#REF! mm						
0 a 7	7.00249	-2.490	± 6.8	±	#REF! mm						
0 a 8	8.00303	-3.030	± 7.6	±	#REF! mm						