

Clase de Exactitud (Accuracy Class): Clase III (Class III)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre
m	m	mm	mm	(Uncertainty)	m	m	mm	mm	(Uncertainty)
0 a 0.1	0.09917	0.830	± 1.2	± 3.2E-01 mm					
0 a 0.5	0.49915	0.850	± 1.6	± 3.2E-01 mm					
0 a 1	0.99915	0.850	± 2	± 3.2E-01 mm					
0 a 2	1.99915	0.850	± 2.8	± 3.9E-01 mm					
0 a 3	2.99847	1.530	± 3.6	± 3.9E-01 mm					
0 a 4	3.99787	2.130	± 4.4	± 3.9E-01 mm					
0 a 5	4.99751	2.490	± 5.2	± 3.9E-01 mm					
0 a 6	5.99705	2.950	± 6	± 2.1E-01 mm					
0 a 7	6.99632	3.680	± 6.8	± 2.1E-01 mm					
0 a 8	7.99613	3.870	± 7.6	± 2.1E-01 mm					