

Clase de Exactitud (Accuracy Class): Clase III (Class III)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre
m	m	mm	mm	(Uncertainty)	m	m	mm	mm	(Uncertainty)
0 a 0.1	0.09927	0.730	± 1.2	± 3.2E-01 mm					
0 a 0.5	0.49932	0.680	± 1.6	± 3.2E-01 mm					
0 a 1	0.99935	0.650	± 2	± 3.2E-01 mm					
0 a 2	1.99889	1.110	± 2.8	± 3.9E-01 mm					
0 a 3	2.99847	1.530	± 3.6	± 3.9E-01 mm					
0 a 4	3.99789	2.110	± 4.4	± 3.9E-01 mm					
0 a 5	5.00128	-1.280	± 5.2	± 3.9E-01 mm					
0 a 6	6.00221	-2.210	± 6	± 2.1E-01 mm					
0 a 7	7.00249	-2.490	± 6.8	± 2.1E-01 mm					
0 a 8	8.00303	-3.030	± 7.6	± 2.1E-01 mm					